

- Measles IgG ELISA kit (cat.no. AMEA7330)

The Measles Virus IgG ELISA is intended for the qualitative determination of IgG class antibodies against Measles Virus in human serum or plasma (citrate, heparin).

| | |
|------------------------|--|
| Diagnostic Sensitivity | 97.01% (95% - width of CI: 93.93% .. 98.79%) |
| Diagnostic Specificity | 100.0% (95% - width of CI: 90.0% .. 100.0%) |

- Measles IgM ELISA kit (cat.no. MEAM0330)

The Measles Virus IgM ELISA is intended for the qualitative determination of IgM class antibodies against Measles Virus in human serum or plasma (citrate, heparin).

| | |
|------------------------|--|
| Diagnostic Sensitivity | 100.0% (95% - width of CI: 91.19% .. 100.0%) |
| Diagnostic Specificity | 100.0% (95% - width of CI: 98.71% .. 100.0%) |

- Bartonella ELISA kit (cat.no. BAR0900)

The Bartonella ELISA is intended for the qualitative determination of specific antibodies against Bartonella in human serum or plasma (heparin).

| | |
|------------------------|---|
| Diagnostic Sensitivity | 100.0 % (95% - width of CI: 89.42% .. 100.0%) |
| Diagnostic Specificity | 94.3% (95% - width of CI: 89.13% .. 97.52%) |

- Hepatitis E Virus (HEV) IgG ELISA kit (cat.no. HEVG0780)

The Hepatitis E Virus (HEV) IgG ELISA is intended for the qualitative determination of IgG class antibodies against Hepatitis E Virus (HEV) in human serum or plasma (citrate, heparin).

| | |
|------------------------|---|
| Diagnostic Sensitivity | 100 % (95% confidence interval: 94.94% - 100.0 %) |
| Diagnostic Specificity | 99.36 % (95% confidence interval: 96.5 % - 99.98 %) |

- Hepatitis E Virus (HEV) IgM ELISA kit (cat.no. HEVM0780)

The Hepatitis E Virus (HEV) IgM ELISA is intended for the qualitative determination of IgM class antibodies against Hepatitis E Virus (HEV) in human serum or plasma (citrate, heparin).

| | |
|------------------------|--|
| Diagnostic Sensitivity | 96.15 % (95% confidence interval: 80.36% - 99.9 %) |
| Diagnostic Specificity | 98.59 % (95% confidence interval: 95.94 % - 99.71 %) |

Distributed by:

ImTec

DIAGNOSTICS NV

Contact: info@imtec.be | +32(0)3-2251139